

Ref: 8TMS-L

MEMORANDUM

SUBJECT: Analytical Results--- HAB Emergency Bloom 2016 / 1609026

- FROM: Mark Murphy, Organic Chemist William H. Batschelet, PhD, Laboratory Quality Assurance Officer
- THRU: Mark Burkhardt, PhD, Director Laboratory Services Program
- TO: Tina Laidlaw, 8MO

Clean Water Act

Attached are the analytical results for HAB Emergency Bloom 2016 1609026. The table below shows the number of containers received, the work order number(s) assigned, and the date received:

	1609026	Total
15-Sep-2016	2	2

These samples were prepared, analyzed, and verified by the Region 8 Laboratory according to the requirements of the Laboratory Services Request (LSR) and procedures found in the laboratory Quality Assurance Manual (QSP-001) dated June 16, 2016.

Sample Receipt

All samples were received in acceptable condition except as noted in the Analyst Comments or Appendix A. The number of samples received and analyses are listed in Appendix B.

Sample Analysis

All sample results are reported on an as-received basis except as noted in the Analyst Comments. All samples were analyzed within holding times except as noted in Appendix A. All analyses met QC acceptance criteria except as noted in the Analyst Comments or Appendix A.

Field Measurements

All field measurements met QC acceptance criteria except as noted in the Analyst Comments or Appendix A.

QC Note

Routine sample quality control results such as blanks, matrix spikes, and laboratory duplicates, etc. are reported on the quality control pages of this report. Certain of the reported QC criteria may not be applicable or otherwise affect the data usability. Appendix C summarizes the guidelines used by the Region 8 Laboratory to qualify data. This is a general table and may or may not be applicable to this project.

Analyst Comments

Sample Receipt:

Samples were not collected in required PETG bottles. All samples are qualified since sample preservation requirements were not met.

Microcystins by LC/MS/MS

Station ID: 5	5931231	Date / Tim	e Sampled:	09/13/1	6 09:50	W	orkorder 16	09026	
Comment:	Scofield Reservoir	Matrix: Wa	ater		I	Lab Nu	mber: 16090	026-01	A
Method	Parameter	Results	Units	Qual ifier	. Report Limit		ion or Analyzed	By	Batch
Reg. 8 Lab	Anatoxin-A	< 0.05	ug/L	J	0.05	1	09/15/2016	MAM	1600370
Reg. 8 Lab	Cylindrospermopsin	< 0.05	ug/L	J	0.05	1	09/15/2016	MAM	1600370
Reg. 8 Lab	Microcystin-LR	55.4	ug/L	J	0.50	10	09/15/2016	MAM	1600370
Reg. 8 Lab	Microcystin-RR	0.18	ug/L	J	0.05	1	09/15/2016	MAM	1600370
Reg. 8 Lab	Microcystin-YR	< 0.05	ug/L	J	0.05	1	09/15/2016	MAM	1600370

Station ID: N	Mantua Boat Dock	Date / T	Time Sampled	: 09/12/1	6 18:20	We	orkorder 16	09026	
Comment:	Scofield Reservoir	Matrix:	Water]	Lab Nu	mber: 1609	026-02	A
Method	Parameter	Results	Units	Qual ifier	Report Limit		on or Analyzed	By	Batch
Reg. 8 Lab	Anatoxin-A	< 0.05	ug/L	J	0.05	1	09/15/2016	MAM	1600370
Reg. 8 Lab	Cylindrospermopsin	< 0.05	ug/L	J	0.05	1	09/15/2016	MAM	1600370
Reg. 8 Lab	Microcystin-LR	0.53	ug/L	J	0.05	1	09/15/2016	MAM	1600370
Reg. 8 Lab	Microcystin-RR	< 0.05	ug/L	J	0.05	1	09/15/2016	MAM	1600370
Reg. 8 Lab	Microcystin-YR	< 0.05	ug/L	J	0.05	1	09/15/2016	MAM	1600370

Note: "J" Qualifier indicates an estimated value.

Microcystins by LC/MS/MS - Quality Control

		Reporting		Spike	Source		%REC		RPD
Analyte	Result	Limit	Units	Level	Result	%REC	Limits	RPD	Limit
Batch 1600370 - No Preparation									
Matrix Spike (1600370-MS1)	Sou	ırce: 160902	6-02	Prepared	& Analyze	ed: 09/15/	16		
Anatoxin-A	0.99	0.05	ug/L	1.00	< 0.05	98.7	60-140		
Cylindrospermopsin	0.96	0.05	"	1.00	0.04	92.2	60-140		
Microcystin-LR	1.55	0.05	"	1.00	0.53	102	60-140		
Microcystin-RR	0.88	0.05	"	1.00	0.04	84.1	60-140		
Microcystin-YR	0.90	0.05	"	1.00	0.03	86.6	60-140		
Matrix Spike Dup (1600370-MSD1)	Sou	ırce: 160902	6-02	Prepared	& Analyze	ed: 09/15/	16		
Anatoxin-A	0.88	0.05	ug/L	1.00	< 0.05	88.0	60-140	11.5	30
Cylindrospermopsin	0.89	0.05	"	1.00	0.04	85.3	60-140	7.52	30
Microcystin-LR	1.36	0.05	"	1.00	0.53	83.3	60-140	12.9	30
Microcystin-RR	0.76	0.05	"	1.00	0.04	72.0	60-140	14.7	30
Microcystin-YR	0.77	0.05	"	1.00	0.03	73.9	60-140	15.2	30

NOTE: %REC is percent recovery, Result (less sample contribution) divided by the Spike Level RPD is the Relative Percent Difference (difference between the Result and the Source Result) divided by their average



EPA Region 8 Laboratory 16194 West 45th Drive

Golden, CO 80403

HAB

Sample Submission Form

Instructions: Fill out the station ID, station description, date and time in the chain of custody section (the bottle ID column is for lab use only). Print and sign your name and date the sampler block. Note discrepancies to the sampling/shipping protocols or additional information that may be pertinent in the comments block (backside).

		Chain of Custod	ly		
Waterbody:			LSR#: ((0900	26
Station ID		Station Description	Date	Time	Bottle ID (lab)
713/23/	Gobield 1	les-Mt. view bar ramp	et 9/13/16	6950	
NA	Mantuc	Res	9/12/16	1820	U: [409026-02 P:
					U:
			1.1		P:
					U:
					P:
					U:
					P:
container ner	station, water mat ote any discrepanc ne:	AS and ELISA) (one unpreserver rix. Preservative is 10x sample ies or deviations to this statem Signature:	diluent concentrate	s section.	les are kept office
		Signature:	Date and	time:	Cooler Temp (°C):
Received by I			9/15/16	1	

Page 1 of 2

Project: HAB Emergency Bloom 2016 LSR No: 1609026 Appendix A - Exceptions Report

Preliminary Report

Lab Number	Sample Name	Analysis	Analyte Name	Explanation
1609026-01	5931231	Microcystins by LC/MS/MS_2016	*ALL*	Sample preservation requirements not met
1609026-01RE1	5931231	Microcystins by LC/MS/MS_2016	*ALL*	Sample preservation requirements not met
1609026-02	Mantua Boat Dock	Microcystins by LC/MS/MS_2016	*ALL*	Sample preservation requirements not met

Preliminary Report

Appendix B - Samples and Analysis

Work Order #	<u># Samples</u>	Analysis	Method Name	Lab SOP
1609026	2	Microcystins by LC/MS/MS_2016	Reg. 8 Lab	Draft SOP

QC Check (Symbol)	Flagging Criteria
Initial Calibration (ICAL)	All failing analytes for all samples are qualified as estimated.
Initial Calibration	High failure: All detections for failing analytes for all samples are qualified as estimated.
Verification (ICV) or	Low failure: All failing analytes for all samples are qualified as estimated.
Standard Reference	
Material (SRM)	
Continuing Calibration	High failure: All detections for failing analytes for all associated samples are qualified as
Verification (CCV)	estimated. Low failure: All failing analytes for all associated samples are qualified as
	estimated.
Continuing Calibration	All detections for failing analytes for all associated samples where the concentration in the
Blank (CCB)	blank is greater than 1/10 the amount measured in the sample OR the blank contamination
	otherwise affects the sample results are qualified as estimated.
Blanks (BLK) Preparation	All detections for failing analytes for all samples where the concentration in the blank is
Blank, Method, Trip,	greater than 1/10 the amount measured in the sample OR the blank contamination otherwise
Storage, etc.	affects the sample results are qualified as estimated.
Lab Control Sample	High failure: All detections for failing analytes for all associated samples are qualified as
(LCS) or Standard	estimated. Low failure: All failing analytes for all associated samples are qualified as
Reference Material (SRM)	estimated.
or Blank Spike (BS)	
Matrix Spike (MS)	High failure: All detections for failing analytes in the parent sample are qualified as
	estimated. Low failure: All failing analytes in the parent sample are qualified as estimated.
	No qualification if the native concentration is greater than or equal to 4x the spike
	concentration.
Matrix Spike Duplicate	%R Failure: Same as matrix spike. RPD Failure: All failing analytes in the parent sample are
(MSD)	qualified as estimated.
Duplicate Sample (DUP)	All failing analytes in the parent sample are qualified as estimated. No qualification if the
	native concentration is less than the RL.
Serial Dilution (SD)	All failing analytes in the parent sample are qualified as estimated. No qualification if native
	concentration is less than or equal to 50x the RL.
Detection Limit Standard	High failure: All detections for failing analytes less than or equal to 5x the concentration in
(CRA) or (CRL)	the CRL for all associated samples are qualified as estimated. Low failure: All failing
	analytes less than or equal to 5x the RL for all associated samples are qualified as estimated.
Internal Standard (IS)	All analytes associated with the failing IS are qualified as estimated.
Surrogate Spike (SURR)	High failure: All detections for all analytes associated with the failing surrogate are qualified
	as estimated. Low failure: All analytes associated with the failing surrogate are qualified as
	estimated. If obvious chromatographic interference with the surrogate is present,
	qualification may not be necessary and will be based on the professional judgment of the
	analyst.

Appendix C - Data As	ssessment Guidelines
----------------------	----------------------

Note: The J Qualifier is used to indicate an estimated value.