

**ATTACHMENT 15**

**QUALITY ASSURANCE AND CALIBRATION  
PROCEDURES PLAN FOR  
THC, CO, O<sub>2</sub>, and CO<sub>2</sub> CONTINUOUS EMISSION  
MONITORS**

Attachment 15  
Quality Assurance and Calibration Procedures Plan for  
CO, O<sub>2</sub>, and CO<sub>2</sub> Continuous Emission Monitors

Table of Contents

LIST OF TABLES ..... iii

LIST OF ACRONYMS AND ABBREVIATIONS ..... iv

1. PROJECT DESCRIPTION..... 1

    1.1 Location Description ..... 1

2. RESPONSIBILITIES ..... 1

    2.1 Normal Operation and Maintenance ..... 1

    2.2 Audits and Data Evaluation..... 1

    2.3 Quarterly and Annual Audits ..... 2

    2.4 Training ..... 2

3. INSTRUMENT FUNCTION..... 2

    3.1 Incineration System Description ..... 2

    3.2 Air Pollution Control System..... 2

        3.2.1 Spray Dryer..... 3

        3.2.2 Baghouse..... 3

        3.2.3 Saturator and Wet Scrubber ..... 3

        3.2.4 Induced Draft Fan and Stack ..... 3

    3.3 Stack Gas Continuous Emission Monitoring System (CEMS)..... 3

        3.3.1 Sampling Probes and Gas Conditioning ..... 4

        3.3.2 Gas Conditioning System for CO<sub>2</sub>, O<sub>2</sub>, and CO ..... 4

        3.3.3 Gas Conditioning System for THC..... 5

        3.3.4 Gas Monitor Units ..... 5

        3.3.5 Plant Control System ..... 6

        3.3.6 Data Archiving System..... 6

    3.4 Plant Emission Limitations and Standard ..... 6

4. DATA QUALITY OBJECTIVES ..... 6

    4.1 Summary of Data Quality Criteria ..... 7

        4.1.1 Instrument Zero and Span..... 7

        4.1.2 Calibration Drift..... 7

        4.1.3 Relative Accuracy..... 7

        4.1.4 Response Time..... 7

        4.1.5 Calibration Error ..... 8

        4.1.6 Measurement and Recording Frequency ..... 8

        4.1.7 Hourly Rolling Average Calculation ..... 8

    4.2 Specific Data Quality Objectives for Aragonite CEMS..... 8

5. PERFORMANCE SPECIFICATION TESTING..... 9

    5.1 General Information ..... 9

    5.2 Components of the Performance Specification Test..... 9

        5.2.1 Pretest Preparation ..... 10

        5.2.2 Calibration Drift Test (7-day drift test) ..... 10

5.2.3 Response Time Test.....	10
5.2.4 Calibration Error Test.....	12
5.2.5 Relative Accuracy Test.....	13
5.2.6 Calculations and Equations.....	13
5.2.7 Reporting .....	13
5.3 Performance Specification Testing Details .....	13
6. CALIBRATION PROCEDURES AND FREQUENCY .....	14
6.1 Calibration Procedures .....	15
6.1.1 General Calibration Sequence .....	15
6.1.2 Manual Calibration .....	15
6.2 Instrument Calibration Frequencies .....	15
7. SYSTEM AND COMPONENT AUDITS.....	15
7.1 Daily CEMS Audits .....	16
7.1.1 Daily Inspection Records.....	16
7.2 Quarterly CEMS Audits .....	16
7.3 Annual CEMS Audits.....	16
8. DATA RECORDING, CALCULATIONS, AND REPORTING .....	17
8.1 Records.....	17
8.2 Calculations.....	17
8.3 Reporting.....	17
9. PREVENTIVE MAINTENANCE.....	17
9.1 Daily Preventive Maintenance .....	17
9.2 Preventive Maintenance .....	17
10. CORRECTIVE ACTION .....	18
10.1 System Audits .....	18
10.2 Inspection and Preventive Maintenance.....	19

## LIST OF TABLES

TABLE 4.1	
CEMS DATA QUALITY OBJECTIVES (PERFORMANCE SPECIFICATIONS) .....	8
TABLE 5.1	
PERFORMANCE SPECIFICATION AND REFERENCE METHOD TESTS FOR CEMS AT ARAGONITE .....	13
TABLE 9.1	
MONITOR MAINTENANCE AND RECALIBRATION.....	17
APPENDIX A.....	CEMS CYLINDER GAS RANGES
APPENDIX B .....	CALIBRATION DRIFT DATA SHEET
APPENDIX C .....	CALIBRATION ERROR TEST DATA SHEET
APPENDIX D.....	RESPONSE TIME TEST DATA SHEET
APPENDIX E .....	SEVEN-DAY CALIBRATION DRIFT TEST DATA SHEET

Note: The actual data sheets used may vary in format from those in Appendices B, C, D, and E but will have all the essential information required in these examples.

## LIST OF ACRONYMS AND ABBREVIATIONS

AAO	Air Approval Order
acfm	actual cubic feet per minute
APC	air pollution control
CAA	Clean Air Act
CBO	Control Board Operator
CD	calibration drift
CE	calibration error
CEMS	continuous emission monitoring system
CEM	continuous emission monitor
CO	carbon monoxide
CO <sub>2</sub>	carbon dioxide
DAQ	Utah Division of Air Quality
dscf	dry standard cubic feet
EA	Aragonite Environmental Affairs
EER	Excess Emissions Report
FRP	fiberglass reinforced plastic
ID	induced draft
mA	milli ampere
NIST	National Institute of Standards and Technology
NO <sub>x</sub>	oxides of nitrogen
O <sub>2</sub>	oxygen
PCBs	polychlorinated biphenyls
P&ID	process and instrumentation drawing
ppm	parts per million
PST	performance specification test
QACPP	quality assurance and calibration procedures plan
RA	relative accuracy
RAA	relative accuracy audit
RATA	relative accuracy test audit
RCRA	Resource Conservation and Recovery Act
scfm	standard cubic feet per minute
TSCA	Toxic Substances Control Act
THC	Total Hydrocarbon
UAC	Utah Air Conservation Rules
UDWMRC	Utah Division of Waste Management and Radiation Control

## **1. PROJECT DESCRIPTION**

Clean Harbors Aragonite, LLC operates a hazardous and industrial waste incineration facility located at Aragonite, in Tooele County, Utah. The purpose of this document is to outline the procedures used by Aragonite to comply with RCRA and TSCA requirements specifically for the Total Hydrocarbon Continuous Emission Monitors as required under 40 CFR §60 Appendix B PS8A, for the Carbon Monoxide and Oxygen Continuous Emission Monitors as required under R315-266 Appendix IX §2.1, and for the Carbon Dioxide Continuous Emission Monitors as required under 40 CFR §60 Appendix B PS3. This document is only part of the overall Quality Assurance Plan, required by Utah Division of Air Quality for the entire Continuous Emission Monitoring System (CEMS) at Aragonite. The entire CEMS also includes NO<sub>x</sub> and SO<sub>2</sub> monitors and an annubar to measure stack gas flow.

This Quality Assurance and Calibration Procedures Plan (QACPP) has been developed as Attachment 15 of the State Permit UTD981552177. This plan describes the quality assurance and quality control procedures, including instrument calibration, which will be used to insure that the CEMS data are valid.

The CEMS at Aragonite consist of two complete (redundant) CEM systems. When one of the CEMS is being calibrated, the other CEMS is on-line to insure that the waste feed cutoffs for THC, CO, O<sub>2</sub>, and combustion efficiency are being maintained.

### **1.1 Location Description**

The CEMS analyzers are located in the stack room situated near the base of the stack. Stack gas samples are transported to the instruments via heated, Teflon sampling lines.

## **2. RESPONSIBILITIES**

### **2.1 Normal Operation and Maintenance**

One of the two CEMS will be on-line continuously when the plant is processing waste. Data collection from the CEMS is managed by the plant control system. Daily calibration is initiated by instrumentation technicians or other qualified personnel, in cooperation with the control board operator.

In conjunction with the daily calibration checks, a general inspection of the CEMS is performed. Preventive maintenance is conducted in accordance with Section 9 of this plan. Daily inspections of CEMS are conducted by the Maintenance Department.

### **2.2 Audits and Data Evaluation**

The calibration drift and other calculations are performed by the instrumentation technicians on a daily basis. The Instrumentation Supervisor or other qualified designee shall review the results of the daily calibration and inspection forms. The calibration forms are maintained in the maintenance department or the document vault on site for permanent storage.

### **2.3 Quarterly and Annual Audits**

Aragonite insures that the required quarterly and annual audits are conducted. The actual audit is performed by trained Aragonite personnel or qualified contractors.

### **2.4 Training**

Each department manager is responsible for training employees in their specific job responsibilities. Records of employee training are maintained in the Training Department. Training on this QACPP will be given to all instrument technicians and involved supervisors. Refresher training will be given whenever changes are made to the QACPP.

## **3. INSTRUMENT FUNCTION**

### **3.1 Incineration System Description**

The incinerator facility consists of a horizontal slagging rotary kiln with a vertical afterburner chamber, a gas conditioning and air pollution control train composed of a spray dryer, baghouse, saturator, wet scrubber, induced draft fan, and a stack.

Waste bulk solids are fed into a feed hopper at the kiln front wall and then enter the kiln through the solids feed chute. Containerized wastes are fed to the kiln through the container feed elevator and feed chamber inlet gate. Waste liquids, sludges, and fuels are fed to the kiln through burners or lances at the kiln front wall.

The afterburner chamber has two burners to maintain the minimum required temperatures and provides sufficient volume to hold waste gases for the required residence time.

Gases exit the afterburner chamber into the hot duct. A relief vent is located at the highest elevation of this duct which activates to vent the system under certain plant upset conditions. Under vented conditions there is a net inflow through all unsealed openings. The system is maintained under negative pressure.

### **3.2 Air Pollution Control System**

The Air Pollution Control System (APCS) quenches the exhaust gases to lower temperatures, removes particulate material, and scrubs acid gases from the gas stream.

### **3.2.1 Spray Dryer**

Combustion gases from the afterburner chamber enter the spray dryer. A neutralized scrubbing solution from the downstream scrubbers and make-up water are sprayed into the hot gases, cooling them and evaporating all of the incoming water so that dissolved solids are left as dry crystalline solids.

### **3.2.2 Baghouse**

The gas then flows to the baghouse, where solids are filtered out. As solids build up on the fabric surfaces, the pressure drop across the baghouse increases. The bags are periodically pulsed with compressed air on the discharge side to remove solids, which then fall into the baghouse hoppers.

### **3.2.3 Saturator and Wet Scrubber**

Gas from the baghouse at about 350-520°F travels to the saturator, where a water solution is sprayed into the hot gas to reduce its temperature to less than 225°F and generating a saturated gas stream. An excess of water is used, and the excess is drained to the wet scrubber neutralization tank and recirculated. The saturated gas flows into the wet scrubber which is a two-staged packed bed design.

The majority of the HCl and Cl<sub>2</sub> is removed in the saturator and the first stage of the scrubber, and the majority of the SO<sub>2</sub> is removed in the second stage of the wet scrubber, although a portion of each gas is removed in both scrubbers.

The temperature of the gas stream is further reduced to about 140 to 150°F in the scrubber which causes the majority of the water in the gas stream to condense and results in a much smaller volume of gas.

### **3.2.4 Induced Draft Fan and Stack**

The induced draft (ID) fan creates the draft (partial vacuum) in the incineration train which causes the gases to flow through the system.

The gases are discharged to the atmosphere via a 150 foot high fiberglass reinforced plastic (FRP) stack. The stack is five feet in diameter. Stack instruments include an annubar to measure gas velocity, THC, CO, and CO<sub>2</sub> instruments to measure combustion efficiency, O<sub>2</sub> analyzers, NO<sub>x</sub> and SO<sub>2</sub> analyzers. This QACPP only addresses THC, CO, O<sub>2</sub>, and CO<sub>2</sub> monitors.

## **3.3 Stack Gas Continuous Emission Monitoring System (CEMS)**

The facility has two independent CEM systems. Each consists of the same types of gas monitors. Both of the #1 and #2 CEM systems have separate low and high range CO monitors.



The Continuous Emissions Monitoring System (CEMS) is an extractive system used to measure the stack gas components of total hydrocarbon (THC), carbon monoxide (CO), oxygen (O<sub>2</sub>), and carbon dioxide (CO<sub>2</sub>). The equipment for this system includes a sampling probe with a heated sample line to extract the stack gas, a primary and standby sample conditioning system that cools and dries the gas (CO, CO<sub>2</sub>, and O<sub>2</sub> only), and the analyzers which display the value of each measured gas. Diagrams of the systems are shown in drawings D-800-PI-215 and D-800-PI-216 in Attachment 10.

In the CEM room there are three racks, each containing a complete set of instruments. Each individual instrument measures THC, CO<sub>2</sub>, CO<sub>low</sub>, CO<sub>high</sub>, and O<sub>2</sub>. The CO instrument is dual range; the low range is 0-200 ppm and the high range is 0-10,000 ppm. Switching between ranges is performed by the plant control system. The programming automatically toggles between the low and high range values when the top of the low range is reached. The THC and O<sub>2</sub> instruments are single range instruments with ranges of 0-100 ppm for THC, and 0 – 25% for O<sub>2</sub>.

### **3.3.1 Sampling Probes and Gas Conditioning**

The purpose of these systems is to continuously deliver a representative sample of exhaust gas to the individual gas monitors. For CO, CO<sub>2</sub>, and O<sub>2</sub>, the sample gas is conditioned to remove moisture and particulates and be at a relatively low temperature (near ambient). For THC, the sample temperature must be maintained above dew point and unconditioned except for the removal of particulate matter; no moisture is removed from the sample.

The sample probes for each CEM system have in-line filters. Probes in the stack are located as specified in drawing D-034-PI-212 in Attachment 10. A protective tube is included to prevent cooling of the probe below the dew point of the sample gas.

These probes are used to draw sample gas from the stack. The vacuum to extract the samples is created by vacuum pumps located in the CEMS cabinets. The gas is routed from the stack to the CEM cabinet via a heated sample line.

The sample line consists of sample and calibration tubes and a heating element in an insulated jacket. One tube is used for sample collection. This tube originates at the stack probe and travels to the CEM cabinet. For calibration purposes, a second tube carries certified calibration gasses to the stack probe. A check valve is incorporated to prevent the gas stream from entering the test gas line. The temperature of the sample line is regulated by a temperature controller.

### **3.3.2 Gas Conditioning System for CO<sub>2</sub>, O<sub>2</sub>, and CO**

Upon entering each sample cabinet, the sample gas passes through a pre-cooler which reduces the sample temperature to near ambient. This cooling condenses water vapor and the liquid is separated in a condensate separator and pumped to a drain. The drain pump is controlled by level switches in the condensate separator which provides a contact closure to the local control panel. When the level in the separator rises to the "high" level, the condensate pump starts and

runs until the level decreases to the "low" level. The "high-high" level switch provides an alarm and also stops the sample pump.

The gas sample exits the condensate separator and enters the first path of the refrigerant sample dryer. The dew point of the gas sample is lowered substantially and more water vapor condenses and is removed via a condensate separator.

The gas sample exits the first sample dryer and enters the suction side of the sample pump. The pump elevates the pressure to push the gas sample through the second path of the refrigerant sample dryer which lowers the dew point of the sample to between 32 and 38°F. Vapor is condensed and separated with a condensate separator/pump arrangement.

The gas sample exits the sample dryer and passes through a condensate detector and a sample flow meter. The condensate detector is incorporated to sense any moisture in the sample. If this condition should occur, an alarm light is energized, the sample pump is stopped and the sample gas solenoid valves are closed to prevent any moisture from being sent to the analyzers. The sample flow meter provides a signal to the plant control system to stop all waste feed when the CEM sample flow drops below the set point.

The sample gas is distributed to the analyzers through solenoid valves. After the sample gas passes through the analyzers it is vented to atmosphere.

A remote test gas control panel is attached to each CEM cabinet, and is designed to supply calibration gas to all analyzers in both cabinets. A three position switch is used to switch the high and low range CO span gas to the probe. Since the remote test gas is routed back through the sample line, any gas that is selected is sent to all of the analyzers.

Currently, all calibrations are conducted manually. Provisions have been made for the use of an automatic calibration system, in the event that such a modification is approved and the necessary equipment is purchased. In that case, solenoids are installed which allow both zero and span gas for each instrument to be sent to the sample probe. These solenoids are connected to the same line used with the manual calibration solenoids. These solenoids allow for sequencing the calibration gasses used for all the instruments without interference.

### **3.3.3 Gas Conditioning System for THC**

For THC the gas is unconditioned with the exception of particulate removal. Gas is sent to the analyzer hot, above dew point with no moisture removal.

### **3.3.4 Gas Monitor Units**

The Servomex 4900 is used for monitoring O<sub>2</sub>, CO<sub>low</sub>, CO<sub>high</sub>, and CO<sub>2</sub> in the stack gases. The Thermo Scientific 51i-HT is used for monitoring THC. The gas measurement principle uses the paramagnetic properties of oxygen for measurement. Information for the analyzers in use at Aragonite is given below:

<u>Parameter</u>	<u>Manufacturer</u>	<u>Model No.</u>	<u>Measurement Range</u>
Oxygen #1	Servomex	4900	0 – 25 %
Oxygen #2	Servomex	4900	0 – 25 %
CO <sub>low</sub> #1	Servomex	4900	0 – 200 ppm
CO <sub>high</sub> #1	Servomex	4900	0 – 10,000 ppm
CO <sub>low</sub> #2	Servomex	4900	0 – 200 ppm
CO <sub>high</sub> #2	Servomex	4900	0 – 10,000 ppm
CO <sub>2</sub> #1	Servomex	4900	0 – 20 %
CO <sub>2</sub> #2	Servomex	4900	0 – 20 %
THC #1	Thermo Scientific	51i-HT	0-100 ppm
THC #2	Thermo Scientific	51i-HT	0-100 ppm

The serial numbers for the analyzers currently in service will be tracked in a separate document. This document will be transmitted to the appropriate agencies when the serial number changes.

### 3.3.5 Plant Control System

Data from the #1 and #2 CEM systems are retrieved through the plant control system. A computer interface is provided as part of the control system, and a software package archives the CEMS data.

The plant control system allows for operator interface, alarming, and automatic control. The alarms are set up to provide an indication to the operator of the current status of the on-line (selected) CEMS. Calculated values for diluents and automatic waste feed cutoffs are based on the data from the selected CEMS. The value calculated in the plant control system represents THC and CO corrected to 7% oxygen.

### 3.3.6 Data Archiving System

All data is archived and kept in the plant historian.

## 3.4 Plant Emission Limitations and Standard

The limitations on stack THC and CO emissions at the Aragonite facility are specified in Condition 5.B.4. and 5.D.25. of this permit.

## 4. DATA QUALITY OBJECTIVES

The purpose of the CEMS is to continuously analyze the levels of THC, CO, O<sub>2</sub> and CO<sub>2</sub> in the stack gas to confirm continuous compliance with the permitted emission limits. The minimum data quality objectives for the CEMS are the performance specifications in R315-266 Appendix IX §2.1 and 2.2, and 40 CFR §60 Appendix B PS3, as incorporated in the State Permit.

To achieve these objectives, quality assurance and quality control procedures are used. Quality control consists of all the procedures and activities implemented to control or improve the quality of the data derived from the CEMS. These activities include calibration drift adjustments and audits to determine instrument accuracy and linearity. Quality assurance consists of the review and evaluation of the data and procedures to ensure that the quality control program is working effectively and that the quality objectives are being met.

#### **4.1 Summary of Data Quality Criteria**

The sections below enumerate the specific performance specification criteria found in the relevant permits and regulations. The detailed performance specification requirements are found in R315-266 Appendix IX §2.1 and 2.2, and 40 CFR §60 Appendix B PS3.

##### **4.1.1 Instrument Zero and Span**

The CEMS data measurement range must include zero and a high level or span value. The high level values are selected depending on the magnitude of the emissions of each constituent. The ranges of the constituents monitored by the CEMS were given in Section 3.3.3. The CEMS must allow a determination of calibration drift at the zero and high level values.

##### **4.1.2 Calibration Drift**

As defined in the regulations, calibration drift is the difference between the CEMS readings and the established reference value (certified calibration gas) after a stated period of operation during which no unscheduled maintenance, repair, or adjustment took place. The concentration ranges of the calibration gasses for CO, O<sub>2</sub>, and CO<sub>2</sub> must be between 0 - 20 percent of span for the zero point and between 50 - 90 percent of span for the high level point. The zero calibration gas for THC monitor is 0 - 0.1 percent of span. Appendices A and B present the acceptable concentration ranges for the calibration gases used for the CEMS.

The calibration drift (CD) calculation is used to verify that the CEMS meets the CD quality criteria for determining the emission rates of THC, CO, O<sub>2</sub>, and CO<sub>2</sub>. The CD must be calculated each day.

##### **4.1.3 Relative Accuracy**

Relative accuracy (RA) is a comparison of the emission rate determined by the total CEMS to the same value determined by a Reference Method. The purpose of the RA test is to verify the ability of the total CEMS to provide accurate and representative data. RA audits are conducted at system startup and thereafter during the annual performance specification test period.

##### **4.1.4 Response Time**

Response time is the time interval between the start of a step change in the system input (changing concentration of calibration gas) and the time when the instrument displays 95 percent

of the final value. Response time testing will be conducted annually during the performance specification test period. The performance specifications for THC, CO and O<sub>2</sub> CEMS are given in R315-266, Appendix IX, and also appear in Table 5.1

#### 4.1.5 Calibration Error

Calibration Error (CE) is the difference between the concentration indicated by the CEMS and that of the audit gas (EPA protocol audit gas). The purpose of this test is to verify the accuracy and linearity of the individual THC, CO, and O<sub>2</sub> monitors over the entire measurement range. The CE audit is conducted on a quarterly basis. This test differs from the daily calibrations in that different audit gases are used (EPA protocol vs. certified NIST) and the testing is done at three data points, rather than two. The performance specifications for THC, CO, and O<sub>2</sub> calibration error testing are given in Section 5.2.4 and Table 5.1.

#### 4.1.6 Measurement and Recording Frequency

The sample to be analyzed must pass through the measurement section of each monitor without interruption. The measurement and recording frequency required for THC, CO, O<sub>2</sub>, and CO<sub>2</sub> is addressed in Attachment 16 of the State permit.

#### 4.1.7 Hourly Rolling Average Calculation

The THC and CO rolling averages are calculated using the 60 most recent one-minute averages. The one-minute averages are calculated using a sample taken at least once every 15 seconds. This calculated one hour rolling average is evaluated against the automatic waste feed cutoff setpoints.

### 4.2 Specific Data Quality Objectives for Aragonite CEMS

Specific data quality objectives for the Aragonite CEMS are listed in Table 4.1. These objectives are derived from the applicable regulations and permits. Aragonite performs daily, quarterly, and annual checks to ensure that the data quality objectives are met.

**TABLE 4.1  
CEMS DATA QUALITY OBJECTIVES**

CRITERIA	CO	O <sub>2</sub>	CO <sub>2</sub>	THC
Instrument Zero and Span	0 - 200 ppm (low range) 0 - 10,000 ppm (high range)	0 - 25% by volume	0 - 20% by volume	0 - 100 ppm propane
Calibration Drift (CD)	< 3% of span value for 7 days consecutively	< 0.5% from reference value for 7 days consecutively	< 0.5% from reference value for 7 days consecutively	< 3% of span value for 7 days consecutively

	< 6 ppm (low range) < 300 ppm (high range)	< 0.5% O <sub>2</sub>	< 0.5% CO <sub>2</sub>	< 3 ppm
Relative Accuracy (RA)	≤ 10% mean of RM tests or ≤ 10 ppm of the RM results, whichever is less restrictive.	incorporated into CO RA by requirement for CO corrected to 7% O <sub>2</sub>	< 1% CO <sub>2</sub>	N/A
Response Time	< 2 min	< 2 min	N/A	< 2 min
Calibration Error	≤ 5 percent of span  ≤ 10 ppm (low range) ≤ 500 ppm (high range)	< 0.5% O <sub>2</sub>	N/A	≤ 5% of span  ≤ 5 ppm
Measuring and Recording Frequency	see Attachment 16	see Attachment 16	see Attachment 16	see Attachment 16

## 5. PERFORMANCE SPECIFICATION TESTING

Performance specification testing (PST) refers to the annual quality audits performed on the CEMS. This testing is required by the hazardous waste regulations, and it provides a baseline of reference data for system audits.

### 5.1 General Information

The purpose of the performance specification testing is to assure that the CEMS has been designed, installed, maintained, and operated properly and meets the data quality objectives. The specific data quality requirements and test procedures are given in R315-266 Appendix IX §2.1.6 and 2.2.6, and 40 CFR §60 Appendix B PS3 §6.0. Performance specification testing involves comparing data collected from approximately the same location in the stack using alternate Reference Methods.

Performance specification testing was performed when the CEMS were originally installed, and will be repeated at least annually. Aragonite conducts the performance specification tests using qualified in-house personnel, experienced outside contractors, or both.

### 5.2 Components of the Performance Specification Test

The performance specification testing includes the following required steps: pretest preparation, calibration drift test, response time test, calibration error test, relative accuracy test, calculations, and reporting. These steps are outlined below.

### 5.2.1 Pretest Preparation

To prepare for the PST each CEMS is inspected to verify the operational status of the entire CEMS. Because of daily inspections and maintenance checks throughout the year, it is not expected that significant maintenance or repairs will be required. Also included is preparation for contracted RATA testing. This includes setting up equipment, ensuring needed power connections are available, notification of regulators, and having needed supplies on hand.

### 5.2.2 Calibration Drift Test (7-day drift test)

The procedures for the calibration drift test are specified in R315-266 Appendix IX §2.1.6.1 and 2.2.6.1, and 40 CFR §60 Appendix B PS2 §8.3. During the calibration drift test, the facility is operating at normal conditions. During the seven-day drift test, the calibration drift is determined once each day at approximately 24-hour intervals for seven consecutive days. All of the seven-day high and low CD values must be within the allowable CD limits for THC, CO, O<sub>2</sub>, and CO<sub>2</sub> as specified below:

	<u>CO low</u>	<u>CO high</u>	<u>O<sub>2</sub></u>	<u>CO<sub>2</sub></u>	<u>THC</u>
Calibration			< 0.5%	< 0.5%	
Drift	< 3% span	< 3% span	reference	reference	< 3% span
(24 hours)	(< 6 ppm)	(< 300 ppm)	value	value	(< 3 ppm)
			(< 0.5% O <sub>2</sub> )	(< 0.5% CO <sub>2</sub> )	

The determination of CD is to verify the stability of the monitors over time and to verify that the CEMS conform with the established calibration used for calculating emission rates. The test procedures used are identical to those performed during daily calibrations. The CD test is conducted using both zero level and high level calibration gases. The calibration gases (reference gases) are introduced to the CEMS, and the CEMS response data are recorded. The CD is calculated by determining the difference between the CEMS response and the value of the reference gas. All data are recorded on the calibration drift data sheets. A separate form has been developed for the seven-day calibration drift test (Appendix E). In section 5.3 below, further details of the CD performance specification testing is provided.

### 5.2.3 Response Time Test

A determination of the response time is also conducted during the Performance Specification Test period. The procedures for the response time test are specified in R315-266 Appendix IX §2.1.6.2 and 2.2.6.2.

General Information: This test procedure is required annually for the THC, CO and O<sub>2</sub> monitors. This test is done on the CEMS while they are not on-line. The test gases must pass through the entire CEM system (tubing, gas conditioning, etc.). The test involves measuring the time required for the instruments to respond to a change in concentration. For each monitor, the response time for going from zero concentration to the stack effluent concentration, and from the high or span concentration back down to the stack concentration is measured.

Equipment/Materials Required:

1. Zero and Span gas for THC, CO, and O<sub>2</sub> instruments. The span gasses must be higher than the stack effluent concentration. For the O<sub>2</sub> monitor, air may be used for the high level if the calibration gas concentration is lower than the stack O<sub>2</sub> concentration.
2. Stopwatch or other timing device.
3. Response Time Determination Form (see Appendix D).

Procedure:

1. All measurements are done while CEM is off-line to minimize fluctuations in the stack concentrations of THC, CO, O<sub>2</sub>, and CO<sub>2</sub>.
2. Upscale Measurements-- For each instrument, first switch milliamp (mA) meter to observe the output reading while measuring the stack gas.
  - Allow the instrument to stabilize.
  - The 95% of final stable output reading is determined by:  
95% of stable value (upscale) = [stack output mA - 4.0] x .95 + 4.0  
95% of stable value (downscale) = [stack output mA - 4.0] x 1.05 + 4.0  
(Aragonite CEMS zero scale output current equals 4 mA, 20 mA equals full scale)

Next introduce zero gas or nitrogen, and wait till the instrument has stabilized at the zero reading (no change greater than 1% of full scale (0.16 mA) for 30 seconds).

- Switch the instrument over to sample and measure from the stack.
  - Using a stopwatch, measure the time required for the instrument to reach 95% of the final stable value, as determined above.
  - Repeat this three times for each instrument, and enter the data on the form.
3. Downscale Measurements-- For each instrument, switch over to read stack gas concentrations and determine the mA reading for 95% of the final stable output using the same procedure as for the upscale measurements. Next, introduce high level (span) concentration of gas, and wait until the instruments have stabilized. Switch the instruments over to sample and measure from the



stack. Using a stopwatch, measure the time required for the instrument to reach 95% of the final stable value. Repeat three times for each instrument, and enter the data on the form.

4. For the CO instrument in the upscale measurement, the instrument is going to be changing from zero gas, to essentially zero concentration in the stack. Therefore, the response time is zero. For the CO downscale measurement, measure the time required to reach 5 ppm, rather than zero. For the CO instruments, 4.4 mA is the output for 5 ppm for the 0-200 ppm range, and 4.008 mA is the output for 5 ppm for the 0-10,000 ppm range.

5. For the THC instrument in the upscale measurement, the instrument is going to be changing from zero gas, to essentially zero concentration in the stack. Therefore, the response time is zero. For the THC downscale measurement, measure the time required to reach 0.1 ppm, rather than zero. For the THC instruments, 4.4 mA is the output for < 1 ppm.

6. Calculate the average response times for both the upscale and downscale measurements. The longer of the two averages is the system response time.

### 5.2.4 Calibration Error Test

The procedures for the calibration error test are specified in R315-266 Appendix IX §2.1.6.3 and 2.2.6.3. This test is conducted during the calibration drift test period.

Challenge the THC, CO<sub>low</sub>, CO<sub>high</sub>, O<sub>2</sub>, and CO<sub>2</sub> monitors with zero gas and EPA protocol gas at the following concentrations:

<u>Measurement Point</u>	<u>CO low (ppm)</u>	<u>CO high (ppm)</u>	<u>O<sub>2</sub> (percent)</u>	<u>CO<sub>2</sub> (percent)</u>	<u>THC (ppm)</u>
1	0 - 40	0 - 2000	0 - 2	0 - 2	0 - 0.1
2	60 - 80	3000 - 4000	8 - 10	8 - 10	30 - 40
3	140 - 160	7000 - 8000	14 - 16	14 - 16	70 - 80

(the regulations require 2100 - 2400 ppm for the third point of a CO monitor with a span of 3000 ppm; these values were chosen by Aragonite for use with a 0 - 10,000 ppm instrument range)

Challenge the CEMS at three non-consecutive times at each measurement point as defined above, and record the responses on a Calibration Error Data sheet (see Appendix C). The duration of each gas injection would be sufficient to ensure that the CEMS detectors have fully responded.

All of the CE values must be within the allowable CE limits for THC, CO, O<sub>2</sub>, and CO<sub>2</sub> as specified below:

CO low                      CO high                      O<sub>2</sub>                      CO<sub>2</sub>                      THC

Calibration Error	< 5% span (< 10 ppm)	< 5% span (< 500 ppm)	< 0.5% reference value (< 0.5% O <sub>2</sub> )	< 0.5% reference value (< 0.5% CO <sub>2</sub> )	< 5% span (< 5 ppm)
-------------------	-------------------------	--------------------------	--	---	------------------------

### 5.2.5 Relative Accuracy Test

The procedures for the relative accuracy test are specified in R315-266 Appendix IX §2.1.6.4. During the test, the incinerator is operating at least 50 percent of normal load (total feed rate limit). Relative accuracy is a measure of the difference between the emission rate as determined by the CEMS and that determined using reference methods. The reference method values are determined by sampling from a different port on the stack. Each reference method value is compared against the CEMS value corresponding to the same time under similar conditions (i.e., corrected from dry standard to wet actual conditions).

### 5.2.6 Calculations and Equations

The calculations for CD involve converting milliamp (mA) output readings to concentration and simple arithmetic subtraction. These calculations are specified on the CD drift data sheets (Appendix B).

The calculations for CE, RA, CC (confidence coefficient), standard deviation, and arithmetic mean are given in R315-266 Appendix IX §2.1.7.

### 5.2.7 Reporting

The results of the PST are incorporated into a report upon completion of the tests. This report includes test dates, a description of the procedures used, and a summary of the test results. Appendices to the report include data sheets, calculations, data records, and cylinder gas concentration certifications. The report is provided to State and Federal regulatory agencies 30 days following the completion of the last segment of the PST.

### 5.3 Performance Specification Testing Details

The THC, CO, O<sub>2</sub>, and CO<sub>2</sub> CEMS performance specification testing is repeated annually. The annual performance test is conducted simultaneously with the annual Relative Accuracy Test Audit.

Table 5.1 outlines the PST, Reference Method test, and performance specifications for each constituent monitored by the CEMS at Aragonite. The procedures detailed in R315-266, Appendix IX are incorporated.

**TABLE 5.1  
PERFORMANCE SPECIFICATION TESTS FOR CEMS AT ARAGONITE**

	O <sub>2</sub>	CO, low	CO, high	CO <sub>2</sub>	THC
PST Procedures	R315-266 Appendix IX §2.1	R315-266 Appendix IX §2.1	R315-266 Appendix IX §2.1	40 CFR §60 Appendix B PS3	40 CFR §60 Appendix B PS8A
Calibration Drift (CD) Specification	< 0.5% (vol) from reference value (0.5% (vol) O <sub>2</sub> )	< 3% of span (6 ppm)	< 3% of span (300 ppm)	< 0.5% (vol) from reference value (0.5% (vol) CO <sub>2</sub> )	< 3% of span (3 ppm)
CD Passing Tests	7 consecutive days	7 consecutive days	7 consecutive days	7 consecutive days	7 consecutive days
Relative Accuracy (RA) Specification	Incorporated into CO RA calculation	≤ 10% of mean value of RM test data in units of emission standards, or within 10 ppm of the RM data	≤ 10% of mean value of RM test data in units of emission standards, or within 10 ppm of the RM data	< 1% CO <sub>2</sub>	N/A
RA Passing Tests	Incorporated into CO RA calculation	≥ 9 passing tests, with ≤ 3 failing tests	≥ 9 passing tests, with ≤ 3 failing tests	≥ 9 passing tests, with ≤ 3 failing tests	N/A
Calibration Error (CE) Specification	< 0.5% (vol) from reference value (0.5% (vol) O <sub>2</sub> )	< 5% of span (10 ppm)	< 5% of span (500 ppm)	< 0.5% (vol) from reference value (0.5% (vol) CO <sub>2</sub> )	≤ 5% of span ≤ 5 ppm
Response Time Specification	< 2 minutes	< 2 minutes	< 2 minutes	< 2 minutes	< 2 minutes

## 6. CALIBRATION PROCEDURES AND FREQUENCY

Calibration checks are conducted once each 24 hours on the THC, CO, O<sub>2</sub>, and CO<sub>2</sub> monitors. As long as one CEMS is on line and properly calibrated, the plant remains on waste during CEM calibrations. After the CEMS is calibrated, the control board operator switches to the off-line CEMS to complete the calibration process while the plant continues to process waste.

The calibration drift check data is compared to the calibration drift limits to determine the acceptability of the CEMS calibration results. Calibrations and drift checks may also be performed in conjunction with other test procedures outlined in this QACPP or whenever the CEMS response needs to be verified. The CEMS may also need to be recalibrated after any necessary maintenance procedure (see section 9.3) which could affect the operating condition, or anytime that results indicate the CEMS data may be questionable.

## **6.1 Calibration Procedures**

A basic overview of the calibration procedures is given below.

### **6.1.1 General Calibration Sequence**

The monitors are checked for each constituent against two standard gases of known concentrations, a zero value and a span value. From these checks, the calibration drift (CD) assessments are made.

### **6.1.2 Manual Calibration**

Prior to the daily calibration checks, the instrument technicians will request that the CBO switch to the other CEMS (if the CEMS are fully operational and properly calibrated) or go off waste if the alternate CEMS is not available.

During calibration, the instrument technician directs the flow of calibration gas into the sampling ports in the stack. When this happens, the CEMS are receiving calibration gas rather than stack gas.

Each CEMS response is converted into a concentration value which is compared to the reference calibration gas values. If the calibration drift exceeds the operating limits that appear on the calibration forms (Appendix B), a calibration adjustment is performed.

## **6.2 Instrument Calibration Frequencies**

The THC, CO, O<sub>2</sub>, and CO<sub>2</sub> monitors undergo daily calibration drift checks and are checked when questionable data is being transmitted to the plant control system. They are adjusted and recalibrated whenever the drift limits are exceeded.

## **7. SYSTEM AND COMPONENT AUDITS**

To ensure data quality, regular auditing of the CEMS calibration data and inspections is conducted. Aragonite conducts daily, quarterly, and annual audits of the CEMS as required by pertinent regulations. These audit procedures are described in the sections that follow.

## **7.1 Daily CEMS Audits**

The CD data for the THC, CO, O<sub>2</sub>, and CO<sub>2</sub> CEMS are audited on a daily basis, as required by R315-266 Appendix IX §2.1.10 and 2.2.9, and 40 CFR §60 Appendix B PS3. The quality of the CEMS data undergoes a review to check for completeness, accuracy, and whether or not necessary actions were taken.

The calibration drift data sheet is given in Appendix B. The calibration drift data sheet contains instructions as to what actions are to be taken in response to the calibration drift checks. (Note: the calibration forms pertain also to other CEM analyzers regulated under 40 CFR 60, Air Quality regulations).

### **7.1.1 Daily Inspection Records**

The daily audit includes a review of the records of daily inspections of the CEMS. These inspections include the sample transport and interface system (sample cooling system, moisture traps, pre-cooler fan, cooling system temperature). The daily audit forms include a checklist to confirm that all required inspections have been completed and space to identify any necessary or recommended action that results from the inspections.

The plant control system generates alarm signals resulting from out of range monitor readings and low sample flow rate.

The data recording function is handled by the plant control system and data archiving systems. Bad data from the CEMS will result in an automatic waste feed cutoff (WFCO). Also, the data is viewed by the control board operator (CBO).

## **7.2 Quarterly CEMS Audits**

R315-266 Appendix IX §2.1 and 2.2 require that a calibration error (CE) test be performed at least once each quarter. These regulations also allow a substitution of the Relative Accuracy (RA) test for the CE upon the approval of the Director on a case-by-case basis.

Each quarter, Aragonite performs a CE test for the THC, CO, O<sub>2</sub>, and CO<sub>2</sub> monitors of each CEMS. The CE test procedure checks the accuracy and linearity of these instruments. The pass/fail criteria for the CE test are given in Table 5.1.

## **7.3 Annual CEMS Audits**

The annual CEMS audit consists of the required Performance Specification Testing for THC, CO, O<sub>2</sub>, and CO<sub>2</sub>. The annual PST consists of the Calibration Drift, RATA, Response Time test, and quarterly Calibration Error test, as discussed in section 5.

Utah Air Quality regulations require 45 days advance notice to schedule a RATA test.

## **8. DATA RECORDING, CALCULATIONS, AND REPORTING**

### **8.1 Records**

As explained in section 3, the data from both CEMS is retrieved from the plant control system by the data archiving system. The data is archived electronically. Raw data from each individual CEM as well as calculated values determined from the selected CEMS are archived.

Calibration Data sheets and Cylinder Gas certificates of analysis are kept on site for a minimum of three years.

### **8.2 Calculations**

The Plant Control System performs the necessary calculations to convert the CEMS data into emission rates and THC or CO and combustion efficiency waste feed cutoffs. All of the calculations are in compliance with the requirements of R315-266 Appendix IX §2.1.7, 2.2.7, and 40 CFR §60 Appendix B PS3. The calculations are given in R315-266 Appendix IX §2.1.7 and 2.2.7.

### **8.3 Reporting**

The results of the quarterly calibration error (CE) audits are provided as required. The reports are kept on file at the Aragonite facility and are submitted to TSCA 30 days following completion of the audit.

The results of the annual performance specification testing (PST), which will include RATA, Response Time, Calibration Error, and Calibration Drift test are also provided.

## **9. PREVENTIVE MAINTENANCE**

### **9.1 Daily Preventive Maintenance**

The CEMS at Aragonite are calibrated manually each day. During the daily calibration and inspections, upkeep of the analyzer systems and associated equipment is accomplished. If the required maintenance is more involved or not immediately required, the work is scheduled for a later time. Because of the presence of two independent CEMS, much of the maintenance work can be completed while the plant is processing waste.

### **9.2 Preventive Maintenance**

Maintenance checks are performed in accordance with manufacturer recommendations. For any major maintenance procedures referenced in Table 9.1, the affected monitors must pass all of the

annual performance specification test (PST) procedures, with the exception of the RATA, before being returned to service.

**TABLE 9.1  
MONITOR MAINTENANCE AND RECALIBRATION GUIDELINES**

Maintenance Operation	Recalibration or Recertification Requirements
<u>Minor Maintenance:</u> <ul style="list-style-type: none"> <li>•cleaning / replacing filter elements</li> <li>•tubing replacement</li> <li>•minor adjustments</li> </ul>	Perform Zero and Span calibration checks. If necessary recalibrate in accordance with Appendix C.
<u>Major Maintenance Operations:</u> <ul style="list-style-type: none"> <li>•replacement of circuit boards, amplifiers, or other components used in signal processing.</li> <li>•replacement of detectors or measurement cells</li> <li>•repairs performed by the factory</li> <li>•replacement of an analyzer with either an identical unit or one having the same operating principle.</li> </ul>	Perform all of the performance specification test procedures outlined in section 5 of this plan <i>except</i> for the relative accuracy test audit (RATA).
<u>Modification of Critical Components:</u> <ul style="list-style-type: none"> <li>•replacement of analyzers with units employing a different measurement principal.</li> <li>•replacement of sample probes with different type.</li> </ul>	Perform complete set of performance test procedures, in accordance with section 5.

## 10. CORRECTIVE ACTION

The calibration checks, audits, and inspections outlined in this plan are designed to evaluate whether each CEMS is operating within the allowable limits as defined in the relevant regulations and permits.

### 10.1 System Audits

Generated data, calibration checks, or inspections may occasionally show that the CEMS are not operating within the allowable limits. Any CEMS equipment or components that are found to be operating improperly shall be adjusted, repaired, or replaced. The equipment is not to be used unless calibrations are successfully performed.

If the CEMS data are determined to be invalid, operation is switched to the backup CEMS. If the backup CEMS is also not working properly, waste feed to the incinerator ceases.

If problems become apparent during the system or performance audits, corrective action is implemented.

## **10.2 Inspection and Preventive Maintenance**

Problems identified during inspections and preventive activities are subject to corrective actions. This may involve equipment adjustments, cleaning, equipment repair, or equipment replacement.

DRAFT



**APPENDIX A**

**CEMS CYLINDER GAS RANGES**

DRAFT

### CYLINDER GAS RANGES FOR ARAGONITE CEMS

	CO low	CO high	O <sub>2</sub>	CO <sub>2</sub>	THC
Daily Calibration, Zero	Nitrogen, Industrial Grade or Plant Nitrogen (liq.) 99.99% min. purity	Nitrogen, Industrial Grade or Plant Nitrogen (liq.) 99.99% min. purity	Nitrogen, Industrial Grade or Plant Nitrogen (liq.) 99.99% min. purity	Nitrogen, Industrial Grade or Plant Nitrogen (liq.) 99.99% min. purity	Nitrogen, Industrial Grade or Plant Nitrogen (liq.) 99.99% min. purity
Daily Calibration, Span	100 - 180 ppm CO in N <sub>2</sub> certified traceable to NIST	5000 - 9000 ppm CO in N <sub>2</sub> certified traceable to NIST	12.5 - 22.5% O <sub>2</sub> in N <sub>2</sub> certified traceable to NIST	10 - 18% CO <sub>2</sub> in N <sub>2</sub> certified traceable to NIST	80-90 ppm
Response Time Test	daily calibration gas	daily calibration gas	daily calibration gas	daily calibration gas	daily calibration gas
Calibration Drift Test	daily calibration gas	daily calibration gas	daily calibration gas	daily calibration gas	daily calibration gas
Calibration Error Test point 1	0 - 40 ppm CO in N <sub>2</sub> EPA Protocol Gas	0 - 2000 ppm CO in N <sub>2</sub> EPA Protocol Gas	0 - 2% O <sub>2</sub> in N <sub>2</sub> EPA Protocol Gas	0 - 2% CO <sub>2</sub> in N <sub>2</sub> EPA Protocol Gas	0-0.1 ppm
Calibration Error Test point 2	60 - 80 ppm CO in N <sub>2</sub> EPA Protocol Gas	3000 - 4000 ppm CO in N <sub>2</sub> EPA Protocol Gas	8 - 10% O <sub>2</sub> in N <sub>2</sub> EPA Protocol Gas	8 - 10% CO <sub>2</sub> in N <sub>2</sub> EPA Protocol Gas	30-40 ppm
Calibration Error Test point 3	140 - 160 ppm CO in N <sub>2</sub> EPA Protocol Gas	7000 - 8000 ppm CO in N <sub>2</sub> EPA Protocol Gas (1)	14 - 16% O <sub>2</sub> in N <sub>2</sub> EPA Protocol Gas	14 - 16% CO <sub>2</sub> in N <sub>2</sub> EPA Protocol Gas	70-80 ppm

Note (1): The Aragonite CO monitor has range of 0-10,000 ppm. The regulations call for 2100-2400 ppm for a 0-3000 ppm CO monitor.

**APPENDIX B**

**CALIBRATION DRIFT DATA SHEET**

DRAFT

**CEM CALIBRATION DRIFT DATA SHEET**

**CEM#** \_\_\_\_\_ **DATE:** \_\_\_\_\_

Signature: \_\_\_\_\_  
 Start Time: \_\_\_\_\_  
 End Time: \_\_\_\_\_

Reviewed by: _____
Date: _____

	Calibration Gas Concentration	mA Reading (Zero)	% or ppm (Zero)	mA Reading (Span)	% or ppm (Span)	Instrument Drift (Zero)	Instrument Drift (Span)	mA Reading (Zero) after calibration	% or ppm (Zero) after calibration	mA Reading (Span) after calibration	% or ppm (Span) after calibration
CO <sub>2</sub>											
CO low											
CO high											
O <sub>2</sub>											
NOx											

**Daily Operating Drift Limits**  
 (Recalibrate when exceeded)

CO <sub>2</sub>	0.50%
CO low	6 ppm
CO high	300 ppm
O <sub>2</sub>	0.50%
NOx	50 ppm
THC	3 ppm
SO <sub>2</sub>	25 ppm

**5-day Out-of-Control Drift limits \*\***  
**contact Instrumentation Supervisor**

CO <sub>2</sub>	1.00%
CO low	20 ppm
CO high	1000 ppm
O <sub>2</sub>	1.00%
NOx	50 ppm
THC	3 ppm
SO <sub>2</sub>	25 ppm

**Daily Out-of-Control Drift Limits \*\*\***  
**see note 7.**

CO <sub>2</sub>	2.00%
CO low	40 ppm
CO high	2000 ppm
O <sub>2</sub>	2.00%
NOx	100 ppm
THC	6 ppm
SO <sub>2</sub>	50 ppm

**Instrument scale ranges are:**

CO <sub>2</sub>	0 - 20%
CO low (A)	0 - 200 ppm
CO high (B)	0 - 10000 ppm
O <sub>2</sub>	0 - 25%
NOx	0 - 1000 ppm
THC	0 - 100 ppm
SO <sub>2</sub>	0 - 500 ppm

\*\* instrument is out of control when these limits are exceeded for 5 consecutive days

\*\*\* instrument is out of control when these limits are exceeded

**INSTRUCTIONS**

1. enter cal gas concentrations on form.
2. introduce zero gas and record mA reading, for each instrument.
3. calculate instrument zero response as % (or ppm) = (mA - 4)/16 x instrument scale, and enter value, for each instrument.
4. introduce span gas and record mA reading, for each instrument.
5. calculate instrument span response as % (or ppm) = (mA - 4)/16 x instrument scale, and enter value, for each instrument.
6. recalibrate if the Daily Operating Drift Limits are exceeded, and enter recalibrated zero and span values (mA and % or ppm) on form.
7. notify Maintenance Manager or his / her designee immediately if daily Out-of-Control Drift Limits are exceeded.
8. notify the CBO and shift supervisor immediately if any instrument is unable to meet drift limits, or must remain off-line, with reason and duration.
9. when cal gas bottles are changed out, write the new bottle's information in the comment section below.

**APPENDIX C**

**CALIBRATION ERROR TEST DATA SHEET**

DRAFT

Date of Test:		Test technicians:	
Quarter:			
Monitor Type:			
Manufacturer:			
Serial number:		Span:	

Run Number	Calibration Value	mA Value	Monitor Response	zero/low <sup>1</sup>	mid <sup>1</sup>	high <sup>1</sup>
1 - zero						
2 - mid						
3 - high						
4 - mid						
5 - zero						
6 - high						
7 - zero						
8 - mid						
9 - high						
mean difference =				2	2	2
calibration error, % of span =				3	3	3
				%	%	%

Test times (start and Finish):

Cylinder ID no. \_\_\_\_\_  
 Certification Date: \_\_\_\_\_  
 Expiration Date: \_\_\_\_\_

<sup>1</sup> Absolute Difference = Calibration Value - Monitor Response  
<sup>2</sup> Mean Difference = Sum of Absolute Differences / 3  
<sup>3</sup> Calibration Error = (Mean Difference / Span) x 100%

Cylinder ID no. \_\_\_\_\_  
 Certification Date: \_\_\_\_\_  
 Expiration Date: \_\_\_\_\_

CO	O <sub>2</sub> and CO <sub>2</sub>	THC
<b>pass/fail criteria</b>	<b>pass/fail criteria</b>	<b>pass/fail criteria</b>
< 5% of span	< 0.5% O <sub>2</sub>	≤ 5% of span
< 10 ppm (low)	< 0.5% CO <sub>2</sub>	≤ 5 ppm
< 500 ppm (high)		

Cylinder ID no. \_\_\_\_\_  
 Certification Date: \_\_\_\_\_  
 Expiration Date: \_\_\_\_\_

Comments:

**APPENDIX D**

**RESPONSE TIME TEST DATA SHEET**

DRAFT

**CONTINUOUS EMISSIONS MONITORING SYSTEM  
RESPONSE TIME DETERMINATION**

CEM no. \_\_\_\_\_

Clean Harbors  
Aragonite, LLC

DATE: \_\_\_\_\_

OPERATORS: \_\_\_\_\_

Instrument Measurement Parameter	Manufacturer / Model no.	Serial no.	Stack Conc.	Stack Conc. (mA)	Target (mA) (see note)	Response Time 1 (seconds)	Response Time 2 (seconds)	Response Time 3 (seconds)	Average Response Time (seconds)
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						
			ZERO TO STACK CONCENTRATION						
			HIGH TO STACK CONCENTRATION						

(The longer of the two averages is the system response time)

Target mA (upscale) = [stack mA - 4.0] x .95 + 4.0

Target mA (downscale) = [stack mA - 4.0] x 1.05 + 4.0



**APPENDIX E**

**SEVEN-DAY CALIBRATION DRIFT TEST DATA SHEET**

DRAFT

**CEM 7-DAY CALIBRATION DRIFT DATA SHEET**

**CEM #** \_\_\_\_\_

**Start Date:** \_\_\_\_\_

	Date	CO <sub>2</sub>	CO (low)	CO (high)	O <sub>2</sub>	NOx	Signature
Day #1							
Day #2							
Day #3							
Day #4							
Day #5							
Day #6							
Day #7							

**INSTRUCTIONS**

1. enter the date of each day of the 7 day drift period.
2. place a check mark in each daily square for each monitor that passes the 7-day drift limits listed on this form.
3. enter the word "fail" in the square for any monitor having a calibration drift greater than the limit listed on this form.
4. if a monitor exceeds the 7-day drift limit, recalibrate and start a new 7-day cal. drift form for the monitor(s) that fail.

**7-Day Calibration Drift Limits**

CO <sub>2</sub>	0.50%
CO low	6 ppm
CO high	300 ppm
O <sub>2</sub>	0.50%
NOx	25 ppm
THC	3 ppm

Monitor	Manufacturer	Serial No.
CO <sub>2</sub>		
CO low		
CO high		
O <sub>2</sub>		
NOx		

DRAFT